TECO DIAGNOSTICS

1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
FINISHED KIT QUALITY CONTROL

PRODUCT NAME: **ALBUMIN** MANUFACTURE DATE: 10/12/2009
CATALOG #: A502-480
LOT #: 19556 EXP. DATE: 6/2012

COMPONENTS:	ALBUMIN REAGENT	LOT#:	19554	EXP. DATE:	06/2012
	ALBUMIN STANDARD	LOT#:	21080	EXP. DATE:	06/2012

BLANK: 0.202 AT 600 NM

TOLERANCE LEVELS OF SERUM CONTROLS

SERA ID	EXP. DATE	TOLERANCE RANGE (On TC-84)	TARGET (g/dl)
15271	08/2011	3.2 – 4.2	3.7
15272	08/2011	1.7 – 2.7	2.2

TEST VALUES/RESULTS

SERA ID	1	2	MEAN	# PASS
15271	3.9	3.9	3.9	2/2
15272	2.7	2.8	2.8	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: QA/QC DATE: 10/12/2009