



**CERTIFICATE OF ANALYSIS**  
**FINISHED KIT QUALITY CONTROL**

PRODUCT NAME: ALBUMIN                      MANUFACTURE DATE: 02/09/2010  
CATALOG #: A502-480  
LOT #: 24174    EXP. DATE: 09/2012

COMPONENTS: ALBUMIN REAGENT    LOT#: 20748    EXP. DATE: 09/2012  
                  ALBUMIN STANDARD    LOT#: 20749    EXP. DATE: 09/2012

BLANK: 0.374 AT 600 NM

**TOLERANCE LEVELS OF SERUM CONTROLS**

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (On TC-84)	<u>TARGET</u> (g/dl)
15271	03/2012	3.0 – 4.4	3.7
15272	03/2012	1.9 – 2.9	2.4

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	3.7	3.8	3.8	2/2
15272	2.3	2.6	2.5	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED  
BY:

  
Quality Control

DATE: 02/12/2010