TECO DIAGNOSTICS

1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
FINISHED KIT QUALITY CONTROL

PRODUCT NAME:	**ALBUMIN**	MANUFACTURE DATE:	03/01/2011
CATALOG #:	A502-480		
LOT #:	31056	EXP. DATE:	02/2013

COMPONENTS:	ALBUMIN REAGENT	LOT#:	D110103011	EXP. DATE:	11/2013
	ALBUMIN STANDARD	LOT#:	24180	EXP. DATE:	02/2013

BLANK: 0.421 AT 600 NM

TOLERANCE LEVELS OF SERUM CONTROLS

SERA ID	EXP. DATE	TOLERANCE RANGE (On TC-84)	TARGET (g/dl)
15281	09/2012	3.7-4.6	4.1
15272	03/2012	1.9 – 2.9	2.4

TEST VALUES/RESULTS

SERA ID	1	2	MEAN	# PASS
15281	3.9	4.0	4.0	2/2
15272	2.4	2.4	2.4	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: Quality Control DATE: 03/10/2011