TECO DIAGNOSTICS

1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
FINISHED KIT QUALITY CONTROL

PRODUCT NAME:	**ALBUMIN**	MANUFACTURE DATE:	07/08/2014
CATALOG #:	A502-480		
LOT #:	56015	EXP. DATE:	09/2016

COMPONENTS:	ALBUMIN REAGENT	LOT#:	52776	EXP. DATE:	02/2017
	ALBUMIN STANDARD	LOT#:	50506	EXP. DATE:	09/2016

TOLERANCE LEVELS OF SERUM CONTROLS

SERA ID	EXP. DATE	TOLERANCE RANGE (On TC-Matrix)	TARGET (g/dl)
27611	05/2015	4.1 – 5.0	4.6
27602	03/2015	2.7 – 3.3	3.0

TEST VALUES/RESULTS

SERA ID	MEAN	PASS/FAIL
27611	4.1	PASS
27602	2.8	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY: Susan
Susan Tang
Quality Control Manager

DATE: 07/09/2014