

TECO DIAGNOSTICS



1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: ALKALINE PHOSPHATASE (KINETIC) MANUFACTURE DATE: 11/15/2011
CATALOG #: A505-240
LOT #: 35877 EXP. DATE: 04/2013

COMPONENT: ALKALINE PHOSPHATASE REAGENT LOT#: 28466 EXP. DATE: 04/2013

REAGENT BLANK VS. WATER: 0.381 AT 405 NM

TOLERANCE LEVELS OF SERUM CONTROLS


<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (IU/L)
15281	09/2012	89 - 135	112
15282	09/2012	349 - 550	450

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u>#PASS</u>
15281	120	122	121	2/2
15282	471	468	470	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:


Nour Qasqas
Director of Quality Control

DATE: 11/17/2011