

TECO DIAGNOSTICS



1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: ALKALINE PHOSPHATASE (KINETIC) MANUFACTURE DATE: 06/02/2014
CATALOG #: A505-240
LOT #: 55307 EXP. DATE: 08/2015

COMPONENT: ALKALINE PHOSPHATASE REAGENT LOT#: 46222 EXP. DATE: 08/2015

TOLERANCE LEVELS OF SERUM CONTROLS


<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (on TC Matrix)	<u>TARGET</u> (IU/L)
27611	05/2015	81.4 – 122.1	101.7
27602	03/2015	332.2 – 498.4	415.3

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
27611	112.5	PASS
27602	431.1	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY:


Jacqueline Sanchez
Quality Control

DATE: 06/02/2014