



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: ALKALINE PHOSPHATASE (KINETIC) MANUFACTURE DATE: 03/17/2015
 CATALOG #: A505-240
 LOT #: 60052 EXP. DATE: 11/2016

COMPONENT: ALKALINE PHOSPHATASE REAGENT LOT#: 54823 EXP. DATE: 11/2016

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (on TC Matrix)	<u>TARGET</u> (IU/L)
27621	12/2015	86.4 – 129.6	108.0
27602	03/2015	332.2 – 498.4	415.3

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
27621	101.5	PASS
27602	358.2	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY: Jenifer Ohta
 Jenifer Ohta
 Quality Assurance Manager

DATE: 03/18/2015