



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **ALT (SGPT) - KINETIC** MANUFACTURE DATE: 10/21/2008
CATALOG #: A525-400
LOT#: 17078 EXP. DATE: 05/2011

COMPONENT: ALT (SGPT) REAGENT LOT#: 14829 EXP. DATE: 05/2011

REAGENT BLANK VS. WATER: 1.797 AT 340 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET (IU/L)</u>
15221	08/2009	28 – 42	35
15222	08/2009	61 – 93	77

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u>#PASS</u>
15221	35.2	36.5	35.9	2/2
15222	83.5	81.9	82.7	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:


Quality Assurance

DATE: 10/21/2008