



**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **ALT (SGPT) - KINETIC**      MANUFACTURE DATE: 04/15/2010  
 CATALOG #: A525-400  
 LOT#: 25131      EXP. DATE: 04/2013

COMPONENT: ALT (SGPT) REAGENT    LOT#: 24922    EXP. DATE: 04/2013

REAGENT BLANK VS. WATER: 1.829    AT 340 NM

**TOLERANCE LEVELS OF SERUM CONTROLS**

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (IU/L)
15271	3/2012	24.8 – 38.0	31
15272	3/2012	68.4 – 102.6	85.5

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u>#PASS</u>
15271	30.3	30.2	30.3	2/2
15272	89.2	88.5	88.9	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED  
BY:

  
Quality Control

DATE: 04/15/2010