



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **ALT (SGPT) - KINETIC** MANUFACTURE DATE: 10/26/2010
CATALOG #: A525-240
LOT#: 28547 EXP. DATE: 08/2013

COMPONENT: ALT (SGPT) REAGENT LOT#: 26300 EXP. DATE: 08/2013

REAGENT BLANK VS. WATER: 1.870 AT 340 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (IU/L)
15271	03/2012	24.8-38	31
15272	03/2012	69.3-103.8	86.6

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u>#PASS</u>
15271	33.0	33.0	31.5	2/2
15272	85.6	88.3	86.9	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED
BY:


Quality Control

DATE: 10/26/2010