



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **ALT (SGPT) - KINETIC** MANUFACTURE DATE: 02/24/2011
CATALOG #: A525-240
LOT#: 30707 EXP. DATE: 11/2013

COMPONENT: ALT (SGPT) REAGENT LOT#: 29357 EXP. DATE: 11/2013

REAGENT BLANK VS. WATER: 1.744 AT 340 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET (IU/L)</u>
15281	09/2012	28-42	35
15272	03/2012	69.3-103.8	86.6

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u>#PASS</u>
15281	34.1	33.9	34.0	2/2
15272	82.2	85.0	83.6	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED
BY:


Quality Control

DATE: 02/24/2011