



**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: ALT (SGPT) - KINETIC      MANUFACTURE DATE: 10/21/2014  
CATALOG #: A525-400  
LOT#: 57696      EXP. DATE: 09/2017

COMPONENT: ALT (SGPT) REAGENT      LOT#: 56840      EXP. DATE: 09/2017

**TOLERANCE LEVELS OF SERUM CONTROLS**


<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (on TC Matrix)	<u>TARGET</u> (IU/L)
27611	05/2015	26.0 – 39.0	32.5
27602	03/2015	71.6 – 107.4	89.5

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
27611	27.8	PASS
27602	83.5	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY:

  
Jacqueline Sanchez  
Quality Control

DATE:

10/22/2014