



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: ALT (SGPT) - KINETIC MANUFACTURE DATE: 04/28/2015
 CATALOG #: A525-240
 LOT#: 60666 EXP. DATE: 12/2017

COMPONENT: ALT (SGPT) REAGENT LOT#: 57705 EXP. DATE: 12/2017

TOLERANCE LEVELS OF SERUM CONTROLS


<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (on TC Matrix)	<u>TARGET</u> (IU/L)
27621	12/2015	25.2 – 37.9	31.5
27622	12/2015	74.8 – 112.3	93.6

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
27621	30.0	PASS
27622	85.7	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY:


 Jacqueline Sanchez
 Quality Control

DATE:

04/29/2015