

TECO DIAGNOSTICS



1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **AMYLASE - KINETIC** MANUFACTURE DATE: 08/11/2010
CATALOG #: A533-60
LOT #: 27045 EXP. DATE: 06/2013

COMPONENT: AMYLASE REAGENT LOT#: 26676 EXP. DATE: 06/2013

REAGENT BLANK VS. WATER: 0.044 AT 405 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> <u>(IU/L)</u>
15271	03/2012	51.7 – 63.2	57.0
15272	03/2012	268.5 – 328.1	298.0

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	61.4	60.2	60.8	2/2
15272	306.5	304.4	305.45	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED
BY:


Quality Control

DATE: 08/02/2010