

TECO DIAGNOSTICS



1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **AMYLASE - KINETIC** MANUFACTURE DATE: 03/01/2011
CATALOG #: A533-60
LOT #: 30483 EXP. DATE: 11/2013

COMPONENT: AMYLASE REAGENT LOT#: 28552 EXP. DATE: 11/2013

REAGENT BLANK VS. WATER: 0.034 AT 405 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> <u>(IU/L)</u>
15281	09/2012	23.5-80.5	67.0
15272	03/2012	268.5 – 328.1	298.0

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15281	75.9	71.9	73.9	2/2
15272	269.4	287.9	278.7	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:  DATE: 03/02/2011
BY: Quality Control :