



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: AMYLASE - KINETIC MANUFACTURE DATE: 05/09/2014
CATALOG #: A533-60
LOT #: 54824 EXP. DATE: 05/2017

COMPONENT: AMYLASE REAGENT LOT#: 54650 EXP. DATE: 05/2017

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (On TC- Matrix)	<u>TARGET</u> (IU/L)
27611	05/2015	66.1 – 99.2	82.6
27602	03/2015	327.1 – 490.6	408.8

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
27611	89.4	PASS
27602	436.0	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY:

Jacqueline Sanchez
Quality Control

DATE:

05/09/2014