

TECO DIAGNOSTICS



1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.  
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880  
Fax: (714) 463-1169  
[www.tecodiag.com](http://www.tecodiag.com)

**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: AMYLASE - KINETIC      MANUFACTURE DATE: 02/25/2015  
CATALOG #: A533-60  
LOT #: 59714      EXP. DATE: 01/2018

COMPONENT: AMYLASE REAGENT      LOT#: 58870      EXP. DATE: 01/2018

**TOLERANCE LEVELS OF SERUM CONTROLS**

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (On TC- Matrix)	<u>TARGET</u> (IU/L)
27621	12/2015	69.6 – 104.5	87.0
27602	03/2015	327.1 – 490.6	408.8

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
27621	84.1	PASS
27602	482.8	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY: \_\_\_\_\_

  
Roxana Magdaleno  
Quality Control

DATE: \_\_\_\_\_

02/25/2015