

TECO DIAGNOSTICS



1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **AST (SGOT) KINETIC** MANUFACTURE DATE: 04/10/09
CATALOG #: A560-400
LOT #: 19066 EXP. DATE: 07/2011

COMPONENT: AST (SGOT) REAGENT LOT #: 16272 EXP. DATE: 07/2011

REAGENT BLANK VS. WATER: 1.699 AT 340 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u>
15261	08/2011	34 – 44	39
15262	08/2011	158 – 200	179

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15261	38.8	42.1	40	2/2
15262	190.7	187.1	189	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED
BY:

Quality Control

DATE: 04/15/09