

TECO DIAGNOSTICS



1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **AST (SGOT) KINETIC** MANUFACTURE DATE: 04/22/2010
CATALOG #: A560-240
LOT #: 24704 EXP. DATE: 04/2013

COMPONENT: AST (SGOT) REAGENT LOT #: 25033 EXP. DATE: 04/2013

REAGENT BLANK VS. WATER: 1.826 AT 340 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u>
15271	03/2012	26-52	39
15272	03/2012	152-229	191

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15221	34.6	37.1	35.9	2/2
15262	177.1	178.7	177.9	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED
BY:


Quality Control

DATE: 04/22/2010