



**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **AST (SGOT) KINETIC** MANUFACTURE DATE: 6/14/2010  
CATALOG #: A560-240  
LOT #: 26110 EXP. DATE: 4/2013

COMPONENT: AST (SGOT) REAGENT LOT #: 25033 EXP. DATE: 4/2013

REAGENT BLANK VS. WATER: 1.727 AT 340 NM


**TOLERANCE LEVELS OF SERUM CONTROLS**

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET (IU/L)</u>
15271	03/2012	26.4 – 52.0	39.0
15272	03/2012	152.4 – 228.6	190.5

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	35.5	38.8	37.2	2/2
15272	181.0	187.2	184.1	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:	 Quality Control	DATE:	<u>6/14/2010</u>
--------------	--	-------	------------------