

TECO DIAGNOSTICS



1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.  
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880  
Fax: (714) 463-1169  
[www.tecodiag.com](http://www.tecodiag.com)

**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **AST (SGOT) KINETIC**      MANUFACTURE DATE: 03/10/2011  
CATALOG #: A560-240  
LOT #: 30346      EXP. DATE: 03/2014

COMPONENT: AST (SGOT) REAGENT      LOT #: 30384      EXP. DATE: 03/2014

REAGENT BLANK VS. WATER: 0.038      AT 340 NM

**TOLERANCE LEVELS OF SERUM CONTROLS**

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u>
15281	09/2012	34.4-52.0	43.1
15272	03/2012	152.4 – 228.6	190.5

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15281	39.6	40.7	40.2	2/2
15272	156.5	168.3	162.4	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED  
BY:

  
Quality Control

DATE: 03/10/2011