

TECO DIAGNOSTICS



1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.  
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880  
Fax: (714) 463-1169  
[www.tecodiag.com](http://www.tecodiag.com)

**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: AST (SGOT) KINETIC      MANUFACTURE DATE: 05/27/2014  
CATALOG #: A560-240  
LOT #: 55216      EXP. DATE: 11/2015

COMPONENT: AST (SGOT) REAGENT      LOT #: 43636      EXP. DATE: 11/2015

**TOLERANCE LEVELS OF SERUM CONTROLS**

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> <u>(on TC-Matrix)</u>	<u>TARGET</u> <u>(U/L)</u>
27611	05/2013	34.4 - 51.6	43.0
27602	03/2015	160.4 - 240.6	200.5

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
27611	41.4	PASS
27602	198.2	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY:

Jacqueline Sanchez  
Quality Control

DATE:

05/27/2014