

TECO DIAGNOSTICS



1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: AST (SGOT) KINETIC MANUFACTURE DATE: 12/04/2014
CATALOG #: A560-400
LOT #: 58392 EXP. DATE: 09/2017

COMPONENT: AST (SGOT) REAGENT LOT #: 57256 EXP. DATE: 09/2017

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (on TC-Matrix)	<u>TARGET</u> (U/L)
27621	12/2015	31.5 – 47.2	39.4
27602	03/2015	160.4 – 240.6	200.5

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
27621	38.0	PASS
27602	190.6	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY:

Jacqueline Sanchez
Quality Control

DATE:

12/05/2014