**TECO DIAGNOSTICS**

**1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.**
**Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880**
**Fax: (714) 463-1169**
**www.tecodiag.com**

# CERTIFICATE OF ANALYSIS
# FINISHED KIT QUALITY CONTROL

| | | | |
|---|---|---|---|
| PRODUCT NAME: | **BILIRUBIN - DIRECT** | MANUFACTURE DATE: | 02/02/2011 |
| CATALOG #: | B538-480 | | |
| LOT #: | 29538 | EXP. DATE: | 06/2012 |

| | | | | | |
|---|---|---|---|---|---|
| COMPONENTS: | 1. DIRECT BILIRUBIN REAGENT | LOT#: | 30332 | EXP. DATE: | 01/2014 |
| | 2. NITRITE REAGENT | LOT#: | 29410 | EXP. DATE: | 11/2013 |
| | 3. CALIBRATOR (5 mg/dL) | LOT#: | 19780 | EXP. DATE: | 06/2012 |

BLANK: -0.017 AT 550 NM

## TOLERANCE LEVELS OF SERUM CONTROLS

| SERA ID | EXP. DATE | TOLERANCE RANGE (On TC-84) | TARGET (mg/dl) |
|---|---|---|---|
| 15281 | 09/2012 | 0.21– 0.8 | 0.4 |
| 15272 | 03/2012 | 1.5-2.3 | 1.9 |

## TEST VALUES/RESULTS

| SERA ID | 1 | 2 | MEAN | # PASS |
|---|---|---|---|---|
| 15281 | 0.6 | 0.7 | 0.65 | 2/2 |
| 15272 | 1.3 | 1.4 | 1.35 | 2/2 |

COMMENTS: VISUAL CHECK OK.

APPROVED BY: Quality Control DATE: 02/02/2011