CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME:	**BILIRUBIN DIRECT**	MANUFACTURE DATE:	09/09/2011
CATALOG #:	B538-480		
LOT #:	33897	EXP. DATE:	05/2014

COMPONENTS:	DIRECT BILIRUBIN RGT	LOT#:	32716	EXP DATE:	06/2014
	NITRITE REAGENT	LOT#:	33169	EXP DATE:	07/2014
	BILIRUBIN CALIBRATOR	LOT#:	32109	EXP DATE:	05/2014

TOLERANCE LEVELS OF SERUM CONTROLS

SERA ID	EXP DATE	TOLERANCE RANGE	TARGET (mg/dl)
15281	09/2012	0.1 – 0.7	0.4
15282	09/2012	1.3 – 2.0	1.7

TEST VALUES/RESULTS

SERA ID	1	2	MEAN	# PASS
15281	0.4	0.4	0.4	2/2
15282	1.6	1.6	1.6	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:	Quality Control	DATE:	09/09/2011