

TECO DIAGNOSTICS



1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.  
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880  
Fax: (714) 463-1169  
[www.tecodiag.com](http://www.tecodiag.com)

**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **BILIRUBIN DIRECT** MANUFACTURE DATE: 07/14/2014  
CATALOG #: B538-480  
LOT #: 56131 EXP. DATE: 05/2017

COMPONENTS: 1. DIRECT BILIRUBIN REAGENT LOT#: 55353 EXP. DATE: 06/2017  
2. NITRITE REAGENT LOT#: 56009 EXP. DATE: 07/2017  
3. CALIBRATOR (5 mg/dL) LOT#: 55191 EXP. DATE: 05/2017

**TOLERANCE LEVELS OF SERUM CONTROLS**


<u>SERA ID</u>	<u>EXP DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (mg/dl)
27611	05/2015	0.2-0.5	0.3
27602	03/2015	0.8-1.6	1.2

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS /FAIL</u>
27611	0.4	PASS
27602	1.4	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY:

  
Susan Tang  
Quality Control

DATE:

07/15/2014