



**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **BILIRUBIN TOTAL**      MANUFACTURE DATE: 02/25/2011  
 CATALOG #: B576-480  
 LOT #: 30571      EXP. DATE: 06/2012

<u>COMPONENTS:</u>	<u>CATALOG #</u>	<u>LOT#</u>	<u>EXP DATE</u>
1. TOTAL BILIRUBIN REAGENT	B576-A	30740	03/2014
2. NITRITE REAGENT	B538-B	29410	11/2013
3. CALIBRATOR (5.0 mg/dl)	B538-C	19780	06/2012

BLANK: 0.033 AT 550 nm

**TOLERANCE LEVELS OF SERUM CONTROLS**

<u>SERA ID</u>	<u>EXP DATE</u>	<u>TOLERANCE RANGE</u> (On TC-84)	<u>TARGET</u> (mg/dl)
15281	09/2012	0.8-1.7	1.2
15272	03/2012	4.9-7.3	6.1

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15281	1.5	1.7	1.6	2/2
15272	6.5	7.2	6.85	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: *V. Casey*      DATE: 02/28/2011  
Quality Control