

TECO DIAGNOSTICS



1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.  
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880  
Fax: (714) 463-1169  
[www.tecodiag.com](http://www.tecodiag.com)

**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **BILIRUBIN TOTAL** MANUFACTURE DATE: 07/01/2014  
CATALOG #: B576-480  
LOT #: 55381 EXP. DATE: 11/2016

<u>COMPONENTS:</u>	<u>CATALOG #</u>	<u>LOT#</u>	<u>EXP DATE</u>
1. TOTAL BILIRUBIN REAGENT	B576-A	51751	11/2016
2. NITRITE REAGENT	B538-B	55190	05/2017
3. CALIBRATOR (5.0 mg/dl)	B538-C	55191	05/2017

**TOLERANCE LEVELS OF SERUM CONTROLS**

<u>SERA ID</u>	<u>EXP DATE</u>	<u>TOLERANCE RANGE</u> (On TC-84)	<u>TARGET</u> (mg/dl)
27611	05/2015	0.9-1.4	1.1
27602	03/2015	3.9-5.9	4.9

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>MEAN</u>	<u># PASS</u>
27611	1.0	2/2
27602	5.0	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: Susan  
Susan Tang  
Quality Control Manager

DATE: 07/02/2014