



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **UREA NITROGEN (BUN) KINETIC** MANUFACTURE DATE: 6/30/2009
 CATALOG #: B550-240
 LOT #: 19616 EXP. DATE: 08/2011

COMPONENTS: 1. BUN REAGENT LOT#: 18420 EXP. DATE: 08/2011
 2. BUN STANDARD LOT#: 19228 EXP. DATE: 10/2011

REAGENT BLANK VS. WATER: 2.185 AT 340 NM

TOLERANCE LEVELS OF CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (On TC-84)	<u>TARGET</u> (mg/dl)
15261	08/2011	13.3 – 16.9	15.1
15262	08/2011	35.3 – 44.9	40.1

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15261	16.9	16.2	16.55	2/2
15262	43.0	43.1	43.05	2/2

COMMENTS: VISUAL CHECK OK.

APPROVE
D BY:


Quality Control

DATE : 6/30/2009