



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **UREA NITROGEN (BUN) KINETIC** MANUFACTURE DATE: 10/15/2010
CATALOG #: B550-240
LOT #: 28257 EXP. DATE: 06/2012

COMPONENTS: 1. BUN REAGENT LOT#: 22781 EXP. DATE: 06/2012
2. BUN STANDARD LOT#: 24042 EXP. DATE: 12/2012

REAGENT BLANK VS. WATER: 1.717 AT 340 NM

TOLERANCE LEVELS OF CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (On TC-84)	<u>TARGET</u> (mg/dl)
15271	03/2012	11.2-16.8	14.0
15272	03/2012	36.4-54.6	45.5

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	15.9	14.9	15.4	2/2
15272	45.9	45.1	45.5	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED
BY:


Quality Control

DATE: 10/15/2010