



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: UREA NITROGEN (BUN) KINETIC MANUFACTURE DATE: 09/26/2014
 CATALOG #: B550-400
 LOT #: 57345 EXP. DATE: 10/2016

COMPONENTS: 1. BUN REAGENT LOT#: 54285 EXP. DATE: 10/2016
 2. BUN STANDARD LOT#: 54094 EXP. DATE: 10/2016

TOLERANCE LEVELS OF CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (on TC Matrix)	<u>TARGET</u> (mg/dl)
27611	05/2015	14.6 – 18.6	16.6
27602	03/2015	40.5 – 48.5	44.5

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
27611	14.9	PASS
27602	42.8	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY: _____

Jacqueline Sanchez
Quality Control

DATE: _____

09/29/2014