

TECO DIAGNOSTICS

1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
 Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
 Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: UREA NITROGEN (BUN) KINETIC MANUFACTURE DATE: 07/20/2015
 CATALOG #: B550-240
 LOT #: 61788 EXP. DATE: 09/2017

COMPONENTS: 1. BUN REAGENT LOT#: 59096 EXP. DATE: 10/2017
 2. BUN STANDARD LOT#: 59978 EXP. DATE: 09/2017

TOLERANCE LEVELS OF CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (on TC Matrix)	<u>TARGET</u> (mg/dl)
27621	12/2015	14.9 – 18.9	16.9
27622	12/2015	45.1 – 54.1	49.6

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
27621	17.3	PASS
27622	49.6	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY:

Jacqueline Sanchez
 Quality Control

DATE: