



Certificate of Analysis

Product: Creatinine (Jaffé)
Catalogue Number: BXC0111T
Batch Number: 162691
Manufactured Date: 2016/11
Expiry: 2018/07

This product has been tested by Quality Control and has met all final QC specifications. This product is released for packing and is suitable for use until the stated expiry date.

The following results were obtained:

Control Description	Lot Number	Target (U/l)	Range (U/l)	Result (U/l)
Human Serum Normal	HSN026	218	185-251	220.5
Human Serum Elevated	HSE026	449	381-516	446.4

QC PASS/FAIL: **PASS**

Products containing Human Source Material have been tested at the donor level for HbsAg and antibodies to HIV1, HIV2 and HCV and found to be non-reactive.

Signed: *Cherith Morris* (QC Manager) Date: 2016/11

