



Certificate of Analysis

Product: AST (GOT) Liquid Stable
Catalogue Number: BXC0203F
Batch Number: 161697
Manufactured Date: 2016/07
Expiry Date: 2017/12

This product has been tested by Quality Control and has met all final QC specifications.
This product is released for packing and is suitable for use until the stated expiry date.

This product was produced according to ISO 13485 guidelines.

The following results were obtained:

Control Description	Lot Number	Target (U/l)	Range (U/l)	Result (U/l)
Human Serum Normal	HSN023	43	35.8-50.2	41.72
Human Serum Elevated	HSE023	212	176.6-247.4	213.4

QC PASS/FAIL: **PASS**

Products containing Human Source Material have been tested at the donor level for HBsAg and antibodies to HIV1, HIV2 and HCV and found to be non-reactive.

Signed: *Catherine Caskey*

Date: 2016/07