

Certificate of Analysis

Product: Cholesterol CHOD-PAP
 Catalogue Number: BXC0261D
 Batch Number: 161673
 Manufacture: 2016/07
 Expiry: 2018/06

This product has been tested by Quality Control and has met all final QC specifications. This product is released for packing and is suitable for use until the stated expiry date.

This product was produced according to ISO 13485 guidelines.

The following results were obtained:

Control Description	Lot Number	Target (mmol/l)	Range (mmol/l)	Result (mmol/l)
Human Serum Normal	HSN023	6.3	5.25-7.35	6.41
Human Serum Elevated	HSE023	2.6	2.16-3.04	2.76

QC PASS/FAIL: **PASS**

Products containing Human Source Material have been tested at the donor level for HbsAg and antibodies to HIV1, HIV2 and HCV and found to be non-reactive.

Signed: *Catherine Caskey*

Date: 2016/07