



Fortress Diagnostics Limited

Unit 2C  
Antrim Technology Park,  
BT41 1QS  
United Kingdom

Telephone +44 28 94487676  
Facsimile +44 28 94469933

Email [info@fortressdiagnostics.com](mailto:info@fortressdiagnostics.com)  
Website [www.fortressdiagnostics.com](http://www.fortressdiagnostics.com)

## Certificate of Analysis

Product: Triglycerides (Monoliquid)  
Catalogue Number: BXC0271T  
Batch Number: 161670  
Manufactured Date: 2016/07  
Expiry: 2017/11

This product has been tested by Quality Control and has met all final QC specifications. This product is released for packing and is suitable for use until the stated expiry date.

This product was produced according to ISO 13485 & ISO 9001 guidelines.

The following results were obtained:

Control Description	Lot Number	Target (mmol/l)	Range (mmol/l)	Result (mmol/l)
Human Serum Normal	HSN023	2.02	1.68-2.37	2.15
Human Serum Elevated	HSE023	1.03	0.85-1.21	1.01

QC PASS/FAIL: **PASS**

Products containing Human Source Material have been tested at the donor level for HBsAg and antibodies to HIV1, HIV2 and HCV and found to be non-reactive.

Signed: *Catherine Caskey*

Date: 2016/07



ISO 13485:2003  
Cert. No. GB04 / 63491



ISO 9001:2000  
Cert. No. GB04 / 63490