



CERTIFICATE OF ANALYSIS

Cat No: BXC0312A
Lot No: 162763
Manufactured: 2016/11
Expiry: 2019/04
Product: Human Assayed Control Normal

PARAMETER	QUALITY ASSURANCE
Appearance Lyophilised	White fluffy lyophilised plug
Appearance after reconstitution	Slightly turbid solution
Appearance after 6 hours of manufacturing @ 25°C – open tube	Maintains conformity
Moisture Content	<1.0%
Bio-burden	<5 CFU/ml

HOMOGENEITY	QUALITY ASSURANCE (CV <10%)
Alkaline Phosphatase across 30 bottles	1.22 % CV
ALT across 30 bottles	1.42 % CV
AST across 30 bottles	0.66 % CV
Gamma GT	1.22 % CV
Albumin	0.94 % CV
Calcium	2.13 % CV
Total Cholesterol	1.22 % CV

QC PASS/FAIL: PASS

Signed: *Cherith Morris* (QC Manager) **Date:** 2016/09

