



**Fortress Diagnostics Limited**

Unit 2C  
Antrim Technology Park,  
BT41 1QS  
United Kingdom

Telephone +44 28 94487676  
Facsimile +44 28 94469933

Email info@fortressdiagnostics.com  
Website www.fortressdiagnostics.com

### Certificate of Analysis

Product: HDL / LDL Calibrator  
Catalogue Number: BXC0315B  
Batch Number: 161719  
Lot number: 315-01-051  
Manufactured Date: 2016/07  
Expiry: 2018/06

Assigned Values:

Analyte	Method	Target Value	Units
HDL-Cholesterol	Direct	1.27	mmol/L
		49.03	mg/dL
LDL-Cholesterol	Direct	3.40	mmol/L
		131.27	mg/dL

This product has been tested by Quality Control and has met all final QC specifications. This product is released for packing and is suitable for use until the stated expiry date.

This product was produced according to ISO 13485 & ISO 9001 guidelines.

Products containing Human Source Material have been tested at the donor level for HbsAg and antibodies to HIV1, HIV2 and HCV and found to be non-reactive.

Signed: *Catherine Caskey*

Date: 2016/07



ISO 13485:2003  
Cert. No. GB04 / 63491



ISO 9001:2000  
Cert. No. GB04 / 63490