



Fortress Diagnostics Limited

Unit 2C  
Antrim Technology Park,  
BT41 1QS  
United Kingdom

Telephone +44 28 94468786  
Facsimile +44 28 94469933

Email info@Fortressdiagnostics.com  
website www.Fortressdiagnostics.com

## Certificate of Analysis

Product: C-Reactive Protein (Immunoturbidimetric) Calibrator  
Catalogue Number: BXC0324A  
Batch Number: 161724  
Lot Number: 324-06-003  
Manufactured Date: 2016/07  
Expiry: 2018/01

This product has been tested by Quality Control and has met all final QC specifications. This product is released for packing and is suitable for use until the stated expiry date.

This product was produced according to ISO 13485 & ISO 9001 guidelines.

The following results were obtained:

Calibrator Description	Lot Number	Target (mg/l)	Result (mg/l)
CRP (Single Point) Calibrator	324-06-003	280	280

QC PASS/FAIL: **PASS**

The Fortress CRP (Single point) Calibrator is traceable to ERM-DA 472/IFCC.

Products containing Human Source Material have been tested at the donor level for HbsAg and antibodies to HIV1, HIV2 and HCV and found to be non-reactive.

Signed : *Catherine Caskey*

Date : 2016/07



ISO 13485:2003  
Cert. No. GB04 / 63491



ISO 9001:2000  
Cert. No. GB04 / 63490