

## **CERTIFICATE OF ANALYSIS**

| Cat No:       | BXC0326A                   |  |
|---------------|----------------------------|--|
| Product:      | <b>CRP Control Level 1</b> |  |
| Lot no.       | 161693                     |  |
| Manufactured: | 2016/07                    |  |
| Expiry:       | 2017/12                    |  |

| PARAMETER   | QUALITY ASSURANCE                   |  |
|---|-------------------------------------|--|
| Appearance Lyophilised  | N/A                                 |  |
| Appearance Liquid Stable  | Clear solution                      |  |
| Appearance after 6 hours of manufacturing<br>@ 25°C – open tube | No flocculation – maintains clarity |  |
| рН (25°С)   | 7.0                                 |  |
| Bio-burden  | <2 CFU/bottle                       |  |

| HOMOGENEITY           | QUALITY ASSURANCE |  |
|-----------------------|-------------------|--|
| CRP across 30 bottles | 0.51% CV          |  |

| COMPONENTS | VALUE | RANGE     | UNIT |
|------------|-------|-----------|------|
| CRP        | 2.5   | 2.0 - 3.0 | mg/l |

This product was produced according to ISO 13485 & ISO 9001 guidelines.

QC PASS/FAIL: PASS

Signed: Catherine Caskey

Date: 2016/07

CE