

CERTIFICATE OF ANALYSIS

Cat No: BXC0326A
Product: CRP Control Level 1
Lot no. 162782
Manufactured: 2016/11
Expiry: 2017/12

PARAMETER	QUALITY ASSURANCE
Appearance Lyophilised	N/A
Appearance Liquid Stable	Clear solution
Appearance after 6 hours of manufacturing @ 25°C – open tube	No flocculation – maintains clarity
pH (25°C)	7.0
Bio-burden	<2 CFU/bottle

HOMOGENEITY	QUALITY ASSURANCE
CRP across 30 bottles	0.51% CV

COMPONENTS	VALUE	RANGE	UNIT
CRP	2.5	2.0 – 3.0	mg/l

This product was produced according to ISO 13485 & ISO 9001 guidelines.

QC PASS/FAIL: **PASS** Signed: *Cherith Morris* (QC Manager) Date: 2016/11