

Certificate of Analysis

Product: Gamma GT
 Catalogue Number: BXC0362T
 Batch Number: 171354
 Manufactured Date: 2017/06
 Expiry: 2018/08

This product has been tested by Quality Control and has met all final QC specifications. This product is released for packing and is suitable for use until the stated expiry date.

The following results were obtained:

Control Description	Lot Number	Target (U/l)	Range (U/l)	Result (U/l)
Human Serum Level 1	HSN026	53.86	45.78-61.94	54.4
Human Serum Level 2	HSE026	131.26	111.57-150.95	132.3

QC PASS/FAIL: **PASS**

Products containing Human Source Material have been tested at the donor level for HbsAg and antibodies to HIV1, HIV2 and HCV and found to be non-reactive.

Signed: *Cherith Morris* (QC Manager) Date: 2017/06