



## Certificate of Analysis

Product: Direct HDL Cholesterol  
Catalogue Number: BXC0421C  
Batch Number(s): 161702  
Manufacture Date: 2016/07  
Expiry: 2017/09

This product has been tested by Quality Control and has met all final QC specifications. This product is released for packing and is suitable for use until the stated expiry date.

This product was produced according to ISO 13485 guidelines.

The following results were obtained:

Control Description	Lot Number	Target (mmol/l)	Range (mmol/l)	Result (mmol/l)
Lipid Control Level 1	LPCL01	0.81	0.65 – 0.97	0.80
Lipid Control Level 3	LPCE01	1.74	1.39 – 2.09	1.76

QC PASS/FAIL: **PASS**

Products containing Human Source Material have been tested at the donor level for HbsAg and antibodies to HIV1, HIV2 and HCV and found to be non-reactive.

Signed: *Catherine Caskey*

Date: 2016/07