



Certificate of Analysis

Product: Direct HDL Cholesterol
Catalogue Number: BXC0421C
Batch Number(s): 171372
Manufacture Date: 2017/06
Expiry: 2018/06

This product has been tested by Quality Control and has met all final QC specifications. This product is released for packing and is suitable for use until the stated expiry date.

This product was produced according to ISO 13485 guidelines.

The following results were obtained:

Control Description	Lot Number	Target (mmol/l)	Range (mmol/l)	Result (mmol/l)
Fortress Normal	HSN026	2.02	1.72 - 2.32	2.03
Fortress Elevated	HSE026	0.77	0.65 - 0.89	0.75

QC PASS/FAIL: **PASS**

Products containing Human Source Material have been tested at the donor level for HbsAg and antibodies to HIV1, HIV2 and HCV and found to be non-reactive.

Signed: *Cherith Morris* (QC Manager)

Date: 2017/06