

Certificate of Analysis

Product: Ferritin
 Catalogue Number: BXC0441A
 Batch Number: 171533
 Manufacture Date: 2017/06
 Expiry: 2018/08

This product has been tested by Quality Control and has met all final QC specifications. This product is released for packing and is suitable for use until the stated expiry date.

The following results were obtained:

Control Description	Lot Number	Target (ng/ml)	Range (ng/ml)	Result (ng/ml)
Fort Ferr L1	444-01-010	85	72 - 98	78.12
Fort Ferr L2	444-02-010	165	140 - 190	165.7

QC PASS/FAIL: **PASS**

Products containing Human Source Material have been tested at the donor level for HbsAg and antibodies to HIV1, HIV2 and HCV and found to be non-reactive.

Signed: *Cherith Morris* (QC Manager)

Date: 2017/06