

Certificate of Analysis

Product: Microalbumin
 Catalogue Number: BXC0471A
 Batch Number: 162708
 Manufacture Date: 2016/11
 Expiry Date: 2018/06

This product has been tested by Quality Control and has met all final QC specifications. This product is released for packing and is suitable for use until the stated expiry date.

This product was produced according to ISO 13485 guidelines.

The following results were obtained:

Control Description	Lot Number	Target (mg/l)	Range (mg/l)	Result (mg/l)
Urine Control Level 1	UCN002	10	8.5 - 11.5	9.7
Urine Control Level 2	UCE002	54	45 - 63	55.5

QC PASS/FAIL: **PASS**

Products containing Human Source Material have been tested at the donor level for HbsAg and antibodies to HIV1, HIV2 and HCV and found to be non-reactive.

Signed: *Cherith Morris* (QC Manager) Date: 2016/11