



Certificate of Analysis

Product: Lipase (Colorimetric)
Catalogue Number: BXC0511A
Batch Number: 171532
Manufacture Date: 2017/06
Expiry: 2018/11

This product has been tested by Quality Control and has met all final QC specifications. This product is released for packing and is suitable for use until the stated expiry date.

The following results were obtained:

Control Description	Lot Number	Target (U/l)	Range (U/l)	Result (U/l)
Human Serum Normal	HSN026	30.5	25.9-35.1	29.9
Human Serum Elevated	HSE026	75	46.4-62.7	71.8

QC PASS/FAIL: **PASS**

Products containing Human Source Material have been tested at the donor level for HbsAg and antibodies to HIV1, HIV2 and HCV and found to be non-reactive.

Signed: *Cherith Morris* (QC Manager) Date: 2017/06

