



Certificate of Analysis

Product: HbA1c
Catalogue Number: BXC0672B
Batch Number: 161714
Manufactured Date: 2016/07
Expiry: 2017/10

This product has been tested by Quality Control and has met all final QC specifications. This product is released for packing and is suitable for use until the stated expiry date.

This product was produced according to ISO 13485 guidelines.

The following results were obtained:

Control Description	Lot Number	Target (%HbA1c)	Range (%HbA1c)	Result (%HbA1c)
HbA1c Control L1	675-01-003	5.5	4.4 - 6.7	5.6
HbA1c Control L2	675-02-003	9.8	7.8 - 11.7	10.0

QC PASS/FAIL: **PASS**

Products containing Human Source Material have been tested at the donor level for HbsAg and antibodies to HIV1, HIV2 and HCV and found to be non-reactive.

Signed: *Catherine Caskey*

Date: 2016/07