



**Fortress Diagnostics Limited**

Unit 2C  
Antrim Technology Park,  
BT41 1QS  
United Kingdom

Telephone +44 28 94487676  
Facsimile +44 28 94469933

Email info@fortressdiagnostics.com  
website www.fortressdiagnostics.com

## Certificate of Analysis

Product : Immunoglobulin G (IgG)  
Catalogue Number : BXC0721A  
Batch Number : 161668  
Manufactured Date: 2016/07  
Expiry : 2017/10

This product has been tested by Quality Control and has met all final QC specifications. This product is released for packing and is suitable for use until the stated expiry date.

This product was produced according to ISO 13485 & ISO 9001 guidelines.

The following results were obtained:

Control Description	Lot Number	Target (mg/dl)	Range (mg/dl)	Result (mg/dl)
Protein Control L1	PR 001	568	483 - 653	569.2
Protein Control L2	PR 002	1035	880 - 1190	1033.2

QC PASS/FAIL: **PASS**

Products containing Human Source Material have been tested at the donor level for HbsAg and antibodies to HIV1, HIV2 and HCV and found to be non-reactive.

Signed : *Catherine Caskey*

Date : 2016/07



ISO 13485:2003  
Cert. No. GB04 / 63491



ISO 9001:2000  
Cert. No. GB04 / 63490