

Certificate of Analysis

Product: COMPLEMENT C3

Catalogue Number: BXC0851A
Batch Number: 171430
Manufactured Date: 2017/06
Expiry: 2018/12

This product has been tested by Quality Control and has met all final QC specifications. This product is released for packing and is suitable for use until the stated expiry date.

The following results were obtained:

Control Description	Lot Number	Target (mg/dl)	Range (mg/dl)	Result (mg/dl)
Protein Control L1	641-01-023	62.1	49.5-74.4	68.18
Protein Control L2	642-02-023	104	86.6-121.3	106.3

QC PASS/FAIL: PASS

Products containing Human Source Material have been tested at the donor level for HbsAg and antibodies to HIV1, HIV2 and HCV and found to be non-reactive.

Signed: Cherith Morris (QC Manager) Date: 2017/06